

杭州奥泰生物技术股份有限公司 Hangzhou AllTest Biotech Co.,Ltd	文件号 Document No.: ZTC-QC-005-R-003
传染病、心肌、肿瘤类 COA The Infectious Disease、Cardiology、Tumor COA	生效日期 Effective Date: 2018 年 07 月 02 日

Certificate of Analysis

Product Name: SARS-COV-2 Antigen Rapid Test (Nasal Swab)

Catalog No.: INCP-502H

Batch No.: ATNCP21080057

Quantity:2500PCS

Expiry Date: 2022-08

Date of Sampling: 2021-08-21

Date of Analysis: 2021-08-21

Other information: /

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
Functional Performance	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/
---------------	---

Final QC Conclusion:	This batch of product met the QC Criteria.
-----------------------------	--

QC supervisor: Freeman.zheng

Date: 2021.08.21

